

Bharat Coking Coal Limited (A Subsidiary of CIL) Office of the Chief General Manager(MM) Koyla Bhawan: Koyla Nagar Dhanbad FAX NO 0326-2230183.

Ref.No: PUR/ 540132 / Multislice Spiral CT Scanner /CHD/ 08-09/ Dt. 16-04-2010

Pre-Bid meeting for purchase of Multislice CT Scanner for CHD.

BCCL have floated Open tender under reference No; PUR/ 540132 / Multislice Spiral CT Scanner /CHD/ 08-09 with due date of tender opening on 18.05.2010 for purchase of Multislice CT scanner .The NIT terms stipulated for a pre Bid meeting with respective bidders on 16.04 2010 for clarification of NIT terms with no change /deviation in NIT to be permitted.

As stipulated in NTT terms, three firms reported for pre bid meeting on 16.04.2010 with whom pre bid meeting was held in the office of CGM(MM) with name of the prospective bidders as under:-

1.Mr, Angshuman Sinha. Sr, Excutive Sales, M/s Blue Star Limited, Kolkata

2.Mr. Anirban Chakraborty, Assist. Manager Sales, M/s Erbis Engineering Com. Ltd. Kokkata

3.Mr. Sanjoy Chakraborty, Manager Sales , M/s Philips Electronics India Ltd. Kolkata

The clarifications asked by the firms have been replied in the meeting which are as follows:-

SI	Clarification asked by the prospective	Queries replied by BCCL
No.	bidders	representative
1	To change payment terms from existing to Inland letter of credit payable within 21 days against submission of proof of delivery at your bank.	Not acceptable.
2	LD clause from existing to upper limit up to 5% only in stead of 15%.	Not acceptable.
3	EMD to be allowed to be deposited in form of bank guarantee for equivalent amount of Rs. 6.00 lakhs	Not acceptable.
4	Downtime penalty to be clarified.	As per CMC terms of NIT
5	Specify the type of text book	Agreed to incorporate the name of the Author. Two volumes of text book of latest addition authored by Mr., Haaga.
6	Change in the tilt of Gantry to +/- 25 Degree in place of existing +/- 30 degree.	Not acceptable.
7	Change in the Power Output of X ray Generator to 40 KW or higher in place of existing 50 KW or higher.	Not acceptable.

Mired. Califia.

Cnot 1614

frequent 10 pt (c/4/196

16/4110

8	Change in anode heat storage capacity of X ray tube to 4 MHU or higher in place of existing 5 MHU or higher.	Not acceptable.
9	Change in Scan time of Spiral CT for the full 360 degree rotation to 0.8 Sec or less in place of existing less than 0.5 Sec.	Not acceptable.
10	Turnkey job - Specification Measurement regarding Flooring, False Ceiling and Air Conditioning	It was clarified that the measurement of the existing CT Scan Complex is having area of 102 Sq meter (Approx). In respect of flooring/ Ceiling/AC; it was clarified that it will be Vitrified Tiles Flooring, Aluminium False Ceiling and Duct – able Air Conditioning However interested bidders may visit the site for further assessment of the turn key job. Place of installation will be CT Scan Complex at CHD, Dhabbad
11	Whether turn key job will be considered for evaluation of bid?	Evaluation of bid will be made on the basis of all three including Cost of machine, CMC and Turnkey job.
Executive's representing BCCL		Representative of Prospective bidders.

Executive's representing BCCL

- 1. Sri A.K.Sinha CGM (MM)
- 2. . Sri S A Narain GM(MM)
- 3 Dr,S.C.Gupta CMS, CHD
- 4. Dr. TK Karamakar CMO(Store), CHD
- 5. Dr. C.S Gupta HOD(Radilogy)
- 6. Sri H.Prasad Dy, CMM(P)
- 7.Sri S.S.Ghosh FM(P)
- 8.Dr U Biswas MS, (Radiology)
- 9. Sri P.Ranjan MM(P)

- 1. .Mr, Angshuman Sinha
- 2. Mr. Anirban Chakraborty
- Mr.Sanjoy Chakraborty.

The firms informed that they have no further doubts/ questions against this tender.

(Dr. CS Gupta) 4/20 (A K Sinha) (S A Narain) (Dr. SC Gupta) (Dr. TKKarmakar) HOD, Radiology CMS,CHD CMO(Store) CGM(MM) GM(MM)

Dy, CMM(P)

MS(Radiol)

(Dr U Biswas)

(P.Ranjart) MM(P)

Representative of . 1. M/s Blue Star Limited, Kolkata -- Angshum a

2. M/s Erbis Engineering Com. Ltd. Kolkata--

3. M/s Philips Electronics India .Ltd. Kolkata-- Shalles 16/4/10